



CERTIFICATE

Of Conformity
Low Voltage Directive 2014/35/EU

Registration No.: ATGZAHS190516003

Report No.: GZAHS190516003-01

Applicant : Beijing LaserTell Medical Co., Ltd.

Block1, No.12 Jingsheng South 2nd Rd., JQ Science Park,

Tongzhou District, Beijing, China

Product : Beauty Devices

Identification Model No. : AlexMED, AlexMED Pro, DepiMED,

DepiMED Pro, NdMED, ClearMED, FemiMED, VenaMED, EpiMED, EpiMED Pro, TruMED, ExtrMED, iMED, PluShape, PluShape Pro, HydroCoolol, Faceage, 4Deepoo

Trade Mark : LaserTell

Rating : AC230V, 50/60Hz, 15A

Test Standards : EN 60335-1:2012+A11:2014+A13:2017

EN 62233:2008

The certificate of conformity is based on an evaluation of a sample of the above mentioned product. Technical report and documentation are at the applicant's disposal. This is to certify that the tested sample is in conformity with Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. The certificate does not imply assessment of the series-production of the product. The applicant of the certificate is authorized to use this certificate in connection with EU declaration of conformity specified in Article 15 and Annex IV of the Directive.

<u>Jun. 11, 2019</u> Date



Certified by

Terry Tian Manager

CE

The CE Marking may only be used if all relevant and effective EU Directives are complied with

CE

Anbotek (Guangzhou) Compliance Laboratory Limited

Room 508, Building 2, No.232, Kezhu Road, Science City, Economic & Technological Development Area, Guangzhou, Guangdong, China

Tel: (86)020-82575416 Http://www.anbotek.com Fax: (86)020-82575023 Email: service@anbotek.com